



BIOPHARM

Quality Laboratory Solutions and Chemicals

Phone: 870-389-6114
Phone: 800-443-8465
Fax: 870-389-6604
E-mail: support@bphchem.com

187 S. Tilley Rd. .. Hatfield, AR 71945 .. www.bphchem.com

CERTIFICATE OF CONFORMANCE/ CERTIFICATE OF ANALYSIS

Product Number: BS8858

Product Description: Sulfide Standard
1000 ppm (1000 mg/L)
(Sulfide from Sodium Sulfide)

Lot Number: G2924B

Expiration Date*: 7/29/2026

Requirements: 985.0 ppm to 1015 ppm Sulfide

Analysis Results: 1001 ppm +/- 2 ppm

NIST Reference Material: SRM 136, Potassium Dichromate

*Limited Shelf-life Product

Michael Tilly
Certifying Chemist

7-29-2024
Date